



Declaration of Conformity

| | |
|-----------------------------|--|
| Factory: | Laserlinc, Inc. 777 Zapata Dr. Fairborn, OH 45324 USA Tel: 937-318-2440 |
| Product Description: | Ultrasonic and Laser Measurement System |
| Model Numbers: | Ultragaug+e, PTRB, TLAs+er122, TLAs+er122s, TLAs+er130s, TLAs+er160s, TLAs+er1120s, TLAs+er203, TLAs+er 222, TLAs+er230, TLAs+er260, TLAs+er2120, Triton312, Triton330, Triton331, Triton360 |
| Product Category: | Electrical Equipment for measurement, control, and laboratory use. |

We hereby declare that the products conform to the provisions of the Electromagnetic Compatibility (EMC) Directive 2004/108/EC, and the Essential Health and Safety Requirement of the Low-Voltage Directive 2006/95/EC.

EMC Standards:

Emission: EN 61326-1:2006 (IEC 61326-1:2005), Class A
EN 55011:2009 (CISPR 11:2009) Group 1, Class A

Immunity: EN 61326-1:2006 (IEC 61326-1:2005), Class A, Basic Levels

| | |
|--|--|
| EN 61000-4-2:2009 (IEC 61000-4-2:2008) | 2 kV and 4kV Contact Discharge, 8 kV Air Discharge |
| EN 61000-4-3:2006 (IEC 61000-4-3:2008) | 10V/m, 80% AM, 80 MHz-2.7 GHz |
| EN 61000-4-4:2004 (IEC 61000-4-4:2004) | 2 kV Power Lines, 1 kV Signal Lines > 3 m |
| EN 61000-4-5:2006 (IEC 61000-4-5:2005) | 1 kV DM & 2 kV CM Power Lines |
| EN 61000-4-6:2009 (IEC 61000-4-6:2008) | 3V, 80% AM, 150kHz - 80MHz, Power and Signal Lines > 3 m |
| EN 61000-4-11:2004 (IEC 61000-4-11:2004) | Dip: 1 cycle/0%, 10/12 cycles/40%, 25/30 cycles/70% |
| EN 61000-4-11:2004 (IEC 61000-4-11:2004) | Interrupt: 250/300 cycles/0% |

Safety Standards:

| | |
|----------------------|---|
| EN 61010-1:2001Ed:3 | UL Standard for Safety requirements for electrical equipment – for measurement, control and laboratory use. |
| CAN/CSAC22.2#61010-1 | |
| CENELEC EN 61010-1 | |

Laser Safety Standards: Laser Micrometers

IEC 60825-1:2007 (2nd Edition)
Test Procedure: CCA

Environmental Standards: Ultra Gauge+ IP65

IEC 60529-1, Issued: 2000/02/01: Degrees of Protection Provided by Enclosure 2nd Ed.
Clause 5.2 Dust-tight luminaire (IP6X), Clause 6 Jet-proof luminaire (IPX5)
FCC 47CFR 15B cIB

Supplementary Information:


- This product meets the EMC requirements of the United States (FCC Part 15, Class A), and Canada (ICES-001, Group 1, Class A)
- NRTL (Nationally Recognized Testing Laboratory) – Testing performed by: Intertek Testing Services
3933 US Route 11
Cortland, NY 13045 USA

Intertek's Directory of Listed Products - <http://www.intertek.com/directories/>


Date Issued:

August 31, 2012
Laserlinc, Inc.
777 Zapata Dr.
Fairborn, OH 45324 USA

Authorized Signatures:



Jack Weiss, V.P., Production



Dan Dixon, President